

-----  
Mississauga, ON  
Canada  
-----

STATIM 5000 S501R751  
SN 511020J00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000794  
13:27 16/03/2022  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:13  
PRESSURIZING 5:55  
STERILIZING 7:31  
136.4 C 223kPa 7:31  
136.4 C 221kPa 8:01  
135.9 C 216kPa 8:31  
136.4 C 222kPa 9:01  
136.1 C 220kPa 9:31  
136.4 C 217kPa 10:01  
136.0 C 219kPa 10:31  
136.4 C 223kPa 11:01  
136.4 C 223kPa 11:01  
VENTING 11:01  
AIR DRYING 12:00  
CYCLE COMPLETE 13:46  
-----

Digital Signature #  
8ADD4EBA90C4A225  
-----