
Midmark M11 - v1.0.0
Total Cycles: 2619

Name: RITTER 1
Sterilizer ID: V2178198
11/29/2022 11:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.6 F 0.0
2:00 159.3 F 3.2
4:00 184.6 F 7.2
6:00 205.6 F 12.0
8:00 214.9 F 5.3
10:00 228.2 F 10.4
12:00 230.7 F 7.1
14:00 243.2 F 13.1
16:00 242.5 F 11.6
18:00 255.9 F 18.5
20:00 267.7 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.5 F 29.1
2:00 271.4 F 28.4
3:00 272.5 F 29.2
4:00 272.0 F 28.6

Min 271.4 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.5 F 0.0
5:00 217.8 F 0.0
10:00 232.2 F 0.0
15:00 232.3 F 0.0
20:00 226.9 F 0.0
25:00 226.4 F 0.0
30:00 231.1 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 21:19
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:25
CYCLE COMPLETE

11/29/2022 12:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____