
Midmark M11 - v1.0.5
Total Cycles: 4332

Name: RITTER2
Sterilizer ID: V2313228
02/04/2025 01:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	70.4 F	0.0
2:00	75.9 F	0.0
4:00	113.7 F	0.1
6:00	106.5 F	2.3
8:00	145.1 F	5.9
10:00	171.7 F	10.1
12:00	197.9 F	5.6
14:00	211.9 F	5.9
16:00	224.4 F	10.0
18:00	230.9 F	9.3
20:00	237.9 F	10.7
22:00	235.0 F	8.3
24:00	248.5 F	14.9
26:00	253.5 F	16.8

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 251F
CHAMBER PRESS 15.6PSI
** ITEMS NOT STERILE **

02/04/2025 02:16 PM