
Midmark M11 - v1.0.0
Total Cycles: 1068

Name: RITTER 1
Sterilizer ID: V2178198
08/04/2021 06:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.5 F 0.0
2:00 72.0 F 0.0
4:00 78.7 F 0.0
6:00 114.2 F 0.1
8:00 142.2 F 0.2
10:00 167.3 F 0.2
12:00 189.9 F 1.3
14:00 205.5 F 3.9
16:00 215.5 F 7.0
18:00 224.7 F 10.1
20:00 231.6 F 11.4
22:00 233.8 F 9.1
24:00 241.9 F 12.2
26:00 242.1 F 11.4
28:00 238.8 F 9.7
30:00 251.7 F 15.9
32:00 262.7 F 22.2
34:00 272.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.5 F 28.6
2:00 271.7 F 28.2
3:00 272.5 F 28.5
4:00 271.5 F 28.0

Min 271.5 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.9 F 0.0
5:00 221.6 F 0.0
10:00 214.7 F 0.0
15:00 206.8 F 0.0
20:00 201.9 F 0.0
25:00 218.0 F 0.0
30:00 219.4 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 34:22
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:12:32
CYCLE COMPLETE

08/04/2021 07:57 AM
ACCEPTED: X
REJECTED:
Operator: U01