
Midmark M11 - v1.0.0
Total Cycles: 859

Name: RITTER 1
Sterilizer ID: V2178198
06/01/2021 11:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.9 F 0.0
2:00 138.9 F 1.1
4:00 153.8 F 3.5
6:00 178.9 F 6.8
8:00 207.9 F 11.1
10:00 218.0 F 4.4
12:00 227.2 F 9.3
14:00 238.0 F 10.2
16:00 240.0 F 11.6
18:00 238.8 F 9.8
20:00 251.8 F 16.2
22:00 263.1 F 22.7
24:00 272.1 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.2
1:00 272.4 F 29.2
2:00 272.0 F 28.6
3:00 272.6 F 29.3
4:00 271.8 F 28.7

Min 270.9 F 28.1
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.4 F 0.0
5:00 226.5 F 0.0
10:00 229.7 F 0.0
15:00 235.2 F 0.0
20:00 227.3 F 0.0
25:00 216.8 F 0.0
30:00 215.5 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 24:14
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:02:45
CYCLE COMPLETE

06/01/2021 12:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____