
Midmark M11 - v1.0.0
Total Cycles: 4592

Name: RITTER 1
Sterilizer ID: V2178198
05/29/2024 08:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 73.2 F 0.1
4:00 81.0 F 0.2
6:00 113.3 F 0.3
8:00 147.9 F 1.1
10:00 171.9 F 3.8
12:00 184.1 F 6.9
14:00 201.7 F 9.9
16:00 214.7 F 13.1
18:00 220.0 F 5.7
20:00 228.9 F 9.6
22:00 239.1 F 11.4
24:00 238.7 F 10.5
26:00 236.9 F 8.5
28:00 248.1 F 14.2
30:00 258.8 F 20.0
32:00 268.9 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.1
1:00 271.8 F 28.6
2:00 272.3 F 28.6
3:00 271.3 F 28.0
4:00 272.1 F 28.8

Min 271.2 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.3 F 0.0
5:00 207.6 F 0.0
10:00 186.1 F 0.0
15:00 197.6 F 0.0
20:00 208.3 F 0.0
25:00 211.6 F 0.0
30:00 214.3 F 0.0

Drying Complete

FILLING: 0:45
HEATING: 33:05
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:11:31
CYCLE COMPLETE

05/29/2024 09:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____