
Midmark M11 - v1.0.5
Total Cycles: 3303

Name: RITTER2
Sterilizer ID: V2313228
01/16/2024 11:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 116.4 F 0.0
2:00 124.9 F 0.1
4:00 146.7 F 1.6
6:00 162.3 F 4.7
8:00 183.2 F 8.2
10:00 202.7 F 12.2
12:00 216.9 F 4.7
14:00 227.1 F 8.5
16:00 237.2 F 12.8
18:00 234.6 F 9.5
20:00 240.6 F 10.4
22:00 245.3 F 13.4
24:00 256.6 F 19.5
26:00 266.9 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.9 F 29.8
2:00 272.0 F 29.1
3:00 272.1 F 29.6
4:00 272.8 F 29.7

Min 271.6 F 28.9
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 200.7 F 0.0
10:00 205.9 F 0.0
15:00 201.7 F 0.0
20:00 199.8 F 0.0
25:00 197.9 F 0.0
30:00 198.3 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 27:37
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:06:07
CYCLE COMPLETE

01/16/2024 12:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____