

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4506

Name: RITTER2  
Sterilizer ID: V2313228  
04/11/2025 11:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 94.5 F 0.0  
2:00 94.6 F 0.4  
4:00 98.2 F 1.7  
6:00 140.0 F 4.4  
8:00 168.3 F 7.4  
10:00 186.4 F 10.5  
12:00 221.5 F 9.0  
14:00 215.5 F 6.0  
16:00 226.2 F 9.8  
18:00 237.8 F 10.3  
20:00 236.6 F 10.8  
22:00 236.1 F 8.4  
24:00 247.1 F 14.1  
26:00 257.3 F 19.7  
28:00 266.3 F 25.7  
30:00 271.3 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.9  
1:00 272.1 F 30.2  
2:00 272.3 F 30.1  
3:00 272.6 F 30.0  
4:00 272.4 F 29.8

Min 271.7 F 29.7  
Max 272.7 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.2 F 0.0  
5:00 201.8 F 0.0  
10:00 200.1 F 0.0  
15:00 194.3 F 0.0  
20:00 190.5 F 0.0  
25:00 191.6 F 0.0  
30:00 193.6 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 30:27  
STERILIZING: 4:00  
VENTING: 3:53  
DRYING: 30:00  
TOTAL CYCLE: 01:09:15  
CYCLE COMPLETE

04/11/2025 12:46 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_