

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2971

Name: RITTER2  
Sterilizer ID: V2313228  
10/04/2023 05:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.8 F 0.0  
2:00 155.1 F 2.3  
4:00 176.4 F 6.1  
6:00 202.1 F 11.2  
8:00 220.0 F 6.2  
10:00 233.0 F 12.0  
12:00 235.7 F 9.6  
14:00 236.7 F 8.9  
16:00 248.7 F 15.3  
18:00 261.5 F 22.5  
20:00 273.0 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 272.4 F 29.2  
1:00 271.8 F 29.3  
2:00 271.6 F 29.3  
3:00 271.8 F 29.5  
4:00 271.9 F 29.7

Min 271.2 F 29.1  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.5 F 0.0  
5:00 240.8 F 0.0  
10:00 193.8 F 0.0  
15:00 185.3 F 0.0  
20:00 180.9 F 0.0  
25:00 177.7 F 0.0  
30:00 175.5 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 20:05  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:35  
CYCLE COMPLETE

10/04/2023 05:58 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_