
Midmark M11 - v1.0.0
Total Cycles: 699

Name: MIDMARK
Sterilizer ID: V2178198
02/24/2021 01:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.8 F 0.0
2:00 175.0 F 2.8
4:00 181.7 F 7.0
6:00 213.1 F 12.7
8:00 221.6 F 8.3
10:00 237.1 F 10.0
12:00 242.6 F 11.9
14:00 240.2 F 10.4
16:00 254.5 F 17.6
18:00 266.6 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 271.5 F 28.9
2:00 272.7 F 29.4
3:00 271.4 F 28.7
4:00 272.9 F 29.1

Min 271.2 F 28.6
Max 273.1 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.8 F 0.0
5:00 233.8 F 0.0
10:00 212.8 F 0.0
15:00 203.3 F 0.0
20:00 193.2 F 0.0
Dry Cycle Aborted

FILLING: 1:08
HEATING: 19:27
STERILIZING: 4:00
VENTING: 2:54
DRYING: 23:20
TOTAL CYCLE: 00:50:49

02/24/2021 02:48 PM
ACCEPTED: X
REJECTED:
Operator: U01