
Midmark M11 - v1.0.5
Total Cycles: 133

Name: RITTER2
Sterilizer ID: V2313228
06/23/2021 01:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.9 F 0.0
2:00 155.3 F 2.5
4:00 168.3 F 5.8
6:00 188.8 F 9.8
8:00 210.4 F 9.1
10:00 220.8 F 7.9
12:00 232.5 F 12.4
14:00 234.7 F 9.8
16:00 240.7 F 10.0
18:00 246.0 F 14.3
20:00 256.9 F 20.2
22:00 267.0 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.6
1:00 272.5 F 30.5
2:00 272.3 F 30.1
3:00 271.7 F 29.6
4:00 271.6 F 29.9

Min 271.4 F 29.5
Max 272.7 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.4 F 0.0
5:00 196.1 F 0.0
10:00 201.6 F 0.0
15:00 202.7 F 0.0
20:00 256.1 F 0.0
25:00 241.4 F 0.0
30:00 218.2 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 23:36
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:02:44
CYCLE COMPLETE

06/23/2021 02:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____