
Midmark M11 - v1.0.0
Total Cycles: 4450

Name: RITTER 1
Sterilizer ID: V2178198
04/12/2024 08:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.5 F 0.0
2:00 71.2 F 0.2
4:00 76.0 F 1.2
6:00 106.6 F 3.4
8:00 137.0 F 6.1
10:00 167.3 F 9.2
12:00 188.0 F 12.6
14:00 212.0 F 3.2
16:00 221.9 F 6.2
18:00 230.6 F 9.5
20:00 237.8 F 13.2
22:00 233.9 F 8.9
24:00 243.8 F 12.4
26:00 241.5 F 11.5
28:00 253.1 F 17.2
30:00 263.0 F 23.0
32:00 272.4 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.0 F 28.9
2:00 272.4 F 29.0
3:00 272.6 F 29.3
4:00 272.5 F 29.5

Min 271.6 F 28.9
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.2 F 0.0
5:00 219.1 F 0.0
10:00 226.5 F 0.0
15:00 196.3 F 0.0
20:00 187.3 F 0.0
25:00 200.0 F 0.0
30:00 211.3 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 33:05
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:11:24
CYCLE COMPLETE

04/12/2024 09:16 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____