

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1475

Name: RITTER2  
Sterilizer ID: V2313228  
09/12/2022 11:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.3 F 0.0  
2:00 182.3 F 3.5  
4:00 202.9 F 7.8  
6:00 216.5 F 13.2  
8:00 231.6 F 8.7  
10:00 238.7 F 10.7  
12:00 241.5 F 12.0  
14:00 239.9 F 10.7  
16:00 253.4 F 17.4  
18:00 264.9 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.1 F 29.3  
2:00 271.5 F 29.3  
3:00 272.0 F 29.2  
4:00 271.5 F 29.4  
  
Min 271.3 F 28.8  
Max 272.2 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.5 F 0.0  
5:00 226.8 F 0.0  
10:00 234.9 F 0.0  
15:00 235.5 F 0.0  
20:00 228.1 F 0.0  
25:00 225.8 F 0.0  
30:00 227.0 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 19:48  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:58:16  
CYCLE COMPLETE

09/12/2022 12:25 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_