
Midmark M11 - v1.0.0
Total Cycles: 3756

Name: RITTER 1
Sterilizer ID: V2178198
09/12/2023 09:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.8 F 0.0
2:00 183.5 F 1.9
4:00 188.3 F 5.9
6:00 208.5 F 10.7
8:00 218.7 F 4.8
10:00 229.6 F 9.5
12:00 236.1 F 9.5
14:00 241.8 F 11.9
16:00 239.8 F 10.3
18:00 253.0 F 16.8
20:00 264.3 F 23.5
22:00 271.5 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.5 F 28.8
2:00 271.5 F 28.2
3:00 272.7 F 29.0
4:00 272.0 F 28.4

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 224.8 F 0.0
10:00 227.7 F 0.0
15:00 228.4 F 0.0
20:00 229.0 F 0.0
25:00 230.9 F 0.0
30:00 229.6 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 22:01
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:00:52
CYCLE COMPLETE

09/12/2023 10:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____