
Midmark M11 - v1.0.0
Total Cycles: 2593

Name: RITTER 1
Sterilizer ID: V2178198
11/22/2022 10:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.5 F 0.0
2:00 195.7 F 3.9
4:00 195.6 F 8.8
6:00 221.0 F 8.1
8:00 232.0 F 9.7
10:00 237.6 F 8.2
12:00 243.4 F 12.8
14:00 241.9 F 11.2
16:00 255.3 F 18.2
18:00 266.9 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.4 F 28.9
2:00 271.8 F 28.8
3:00 272.3 F 28.8
4:00 271.7 F 28.7

Min 271.5 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.0 F 0.0
5:00 222.4 F 0.0
10:00 231.9 F 0.0
15:00 231.7 F 0.0
20:00 228.7 F 0.0
25:00 227.3 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 19:25
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:43
CYCLE COMPLETE

11/22/2022 11:38 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____