

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3164

Name: RITTER 1  
Sterilizer ID: V2178198  
04/25/2023 03:37 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 207.6 F 0.0  
2:00 201.6 F 0.0  
4:00 221.1 F 5.0  
6:00 235.1 F 10.3  
8:00 235.8 F 7.7  
10:00 242.8 F 11.6  
12:00 246.5 F 13.4  
14:00 246.4 F 13.1  
16:00 260.2 F 20.6  
18:00 272.3 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.9  
1:00 272.2 F 28.2  
2:00 272.5 F 28.7  
3:00 271.7 F 27.9  
4:00 272.7 F 28.6  
  
Min 271.5 F 27.8  
Max 272.7 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.5 F 0.0  
5:00 228.9 F 0.0  
10:00 233.6 F 0.0  
15:00 229.0 F 0.0  
20:00 226.7 F 0.0  
25:00 226.0 F 0.0  
30:00 224.3 F 0.0  
Drying Complete

FILLING: 1:46  
HEATING: 18:26  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:57:28  
CYCLE COMPLETE

04/25/2023 04:34 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_