

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2107

Name: RITTER2  
Sterilizer ID: V2313228  
03/01/2023 10:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.1 F 0.0  
2:00 182.9 F 2.8  
4:00 195.7 F 7.0  
6:00 216.7 F 12.2  
8:00 221.6 F 7.1  
10:00 236.0 F 12.7  
12:00 236.0 F 10.1  
14:00 234.5 F 8.5  
16:00 248.8 F 15.3  
18:00 260.9 F 22.2  
20:00 272.0 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.3  
1:00 272.4 F 30.1  
2:00 271.9 F 29.5  
3:00 271.2 F 29.3  
4:00 272.1 F 30.2

Min 271.1 F 29.1  
Max 272.5 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.5 F 0.0  
5:00 209.6 F 0.0  
10:00 233.7 F 0.0  
15:00 231.3 F 0.0  
20:00 227.9 F 0.0  
25:00 229.0 F 0.0  
30:00 229.8 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 20:35  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:01  
CYCLE COMPLETE

03/01/2023 11:01 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_