
Midmark M11 - v1.0.5
Total Cycles: 3496

Name: RITTER2
Sterilizer ID: V2313228
04/04/2024 08:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.2 F 0.0
2:00 71.6 F 0.0
4:00 82.5 F 0.0
6:00 110.7 F 0.1
8:00 154.2 F 1.1
10:00 178.8 F 4.4
12:00 193.2 F 7.9
14:00 208.7 F 11.8
16:00 216.9 F 4.9
18:00 228.0 F 8.8
20:00 239.0 F 13.3
22:00 238.3 F 10.5
24:00 235.8 F 8.5
26:00 249.4 F 15.3
28:00 260.7 F 21.7
30:00 271.2 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.5 F 29.2
2:00 272.2 F 29.4
3:00 272.0 F 28.9
4:00 272.5 F 29.7

Min 271.5 F 28.6
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 168.2 F 0.0
5:00 206.0 F 0.0
10:00 212.7 F 0.0
15:00 194.8 F 0.0
20:00 191.9 F 0.0
25:00 191.3 F 0.0
30:00 192.3 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 30:39
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:08:53
CYCLE COMPLETE

04/04/2024 09:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____