
Midmark M11 - v1.0.5
Total Cycles: 861

Name: RITTER2
Sterilizer ID: V2313228
03/02/2022 04:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 212.8 F 0.0
2:00 206.9 F 4.5
4:00 211.5 F 10.4
6:00 221.5 F 6.4
8:00 236.6 F 12.7
10:00 238.4 F 11.1
12:00 238.7 F 10.3
14:00 252.7 F 17.4
16:00 264.8 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.3
1:00 272.1 F 30.1
2:00 272.2 F 29.5
3:00 271.6 F 29.6
4:00 272.6 F 29.9

Min 271.4 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.2 F 0.0
5:00 217.9 F 0.0
10:00 229.4 F 0.0
15:00 229.2 F 0.0
20:00 227.8 F 0.0
25:00 230.1 F 0.0
30:00 228.5 F 0.0

Drying Complete

FILLING: 2:49
HEATING: 17:43
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:43
CYCLE COMPLETE

03/02/2022 04:59 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____