

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000450  
10:09 01/02/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:10  
PRESSURIZING 3:13  
STERILIZING 4:24  
136.0 C 230kPa 4:24  
136.2 C 232kPa 4:54  
136.4 C 231kPa 5:24  
136.4 C 233kPa 5:54  
136.4 C 229kPa 6:24  
136.3 C 232kPa 6:54  
136.3 C 233kPa 7:24  
136.3 C 233kPa 7:54  
136.3 C 233kPa 7:54  
VENTING 7:54  
AIR DRYING 8:50  
CYCLE COMPLETE 8:53  
-----

Digital Signature #  
0F40B8B6E815A98B  
-----