
Midmark M11 - v1.0.0
Total Cycles: 2595

Name: RITTER 1
Sterilizer ID: V2178198
11/22/2022 12:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.2 F 0.0
2:00 199.9 F 3.3
4:00 218.9 F 9.3
6:00 230.1 F 6.4
8:00 236.0 F 11.3
10:00 236.0 F 9.0
12:00 240.3 F 9.2
14:00 249.1 F 14.8
16:00 261.9 F 22.0
18:00 272.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.5 F 28.9
2:00 271.7 F 28.6
3:00 272.4 F 28.9
4:00 271.6 F 28.3

Min 271.5 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 210.6 F 0.0
10:00 218.3 F 0.0
15:00 218.9 F 0.0
20:00 218.9 F 0.0
25:00 220.4 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 2:40
HEATING: 18:19
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:13
CYCLE COMPLETE

11/22/2022 01:44 PM
ACCEPTED: X
REJECTED:
Operator: U02