
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 008131
8:47 03/12/2024

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:54
PRESSURIZING 3:55
STERILIZING 5:05
136.5 C 220kPa 5:05
136.2 C 221kPa 5:35
136.3 C 223kPa 6:05
136.4 C 220kPa 6:35
136.4 C 221kPa 7:05
136.4 C 222kPa 7:35
136.0 C 219kPa 8:05
136.4 C 220kPa 8:35
136.4 C 220kPa 8:35
VENTING 8:35
AIR DRYING 9:26
CYCLE COMPLETE 39:26

Digital Signature #
9B94BBACFBBE019E
