
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 000825
14:45 13/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:44
PRESSURIZING 3:45
STERILIZING 4:53
136.5 C 224kPa 4:53
136.3 C 224kPa 5:23
136.4 C 220kPa 5:53
136.4 C 225kPa 6:23
136.4 C 224kPa 6:53
136.4 C 221kPa 7:23
136.4 C 222kPa 7:53
136.4 C 225kPa 8:23
136.4 C 225kPa 8:23
VENTING 8:23
AIR DRYING 9:18
CYCLE COMPLETE 10:02

Digital Signature #
D6A1E1F0F9D1A13C
