
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005930
9:02 10/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 1:56
STERILIZING 2:59
136.5 C 223kPa 2:59
136.0 C 222kPa 3:29
136.3 C 223kPa 3:59
136.4 C 222kPa 4:29
136.4 C 223kPa 4:59
136.4 C 219kPa 5:29
136.2 C 223kPa 5:59
136.4 C 224kPa 6:29
136.4 C 224kPa 6:29
VENTING 6:30
AIR DRYING 7:21
CYCLE COMPLETE 7:30

Digital Signature #
97AEA0367826716C
