
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002446
14:45 23/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:02
PRESSURIZING 2:08
STERILIZING 3:07
136.0 C 220kPa 3:07
136.3 C 223kPa 3:37
136.4 C 224kPa 4:07
136.3 C 223kPa 4:37
136.2 C 223kPa 5:07
136.4 C 225kPa 5:37
136.4 C 225kPa 6:07
136.3 C 225kPa 6:37
136.3 C 225kPa 6:37
VENTING 6:37
AIR DRYING 7:33
CYCLE COMPLETE 37:33

Digital Signature #
E7173C4879DE6FB6
