
Midmark M11 - v1.0.5
Total Cycles: 3312

Name: RITTER2
Sterilizer ID: V2313228
01/19/2024 03:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.1 F 0.0
2:00 186.6 F 4.3
4:00 207.1 F 10.1
6:00 222.4 F 5.6
8:00 239.2 F 12.3
10:00 239.6 F 10.8
12:00 239.4 F 10.3
14:00 255.0 F 18.2
16:00 267.9 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.4 F 29.0
2:00 272.3 F 29.4
3:00 272.3 F 28.9
4:00 272.3 F 29.3

Min 271.5 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.1 F 0.0
5:00 243.5 F 0.0
10:00 206.0 F 0.0
15:00 201.5 F 0.0
20:00 198.1 F 0.0
25:00 196.2 F 0.0
30:00 195.6 F 0.0

Drying Complete

FILLING: 2:11
HEATING: 17:10
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:56:31
CYCLE COMPLETE

01/19/2024 04:41 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____