

-----  
Midmark M11 - v1.0.5  
Total Cycles: 11

Name: RITTER2  
Sterilizer ID: V2313228  
04/22/2021 01:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 149.9 F 0.0  
2:00 149.2 F 2.4  
4:00 164.5 F 6.0  
6:00 188.1 F 11.0  
8:00 214.1 F 5.7  
10:00 223.8 F 11.3  
12:00 232.2 F 8.9  
14:00 237.9 F 10.3  
16:00 246.2 F 14.7  
18:00 259.3 F 21.7  
20:00 270.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.8  
1:00 272.3 F 30.1  
2:00 272.1 F 30.6  
3:00 271.5 F 29.7  
4:00 272.2 F 30.4  
  
Min 271.4 F 29.7  
Max 272.4 F 30.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.4 F 0.0  
5:00 236.3 F 0.0  
10:00 222.0 F 0.0  
15:00 222.3 F 0.0  
20:00 210.2 F 0.0  
25:00 214.5 F 0.0  
30:00 194.2 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 20:43  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:58:51  
CYCLE COMPLETE

04/22/2021 02:13 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_