
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001191
13:18 27/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:40
PRESSURIZING 3:41
STERILIZING 4:58
136.0 C 221kPa 4:58
136.2 C 222kPa 5:29
136.4 C 225kPa 5:59
136.4 C 222kPa 6:29
136.4 C 225kPa 6:59
136.4 C 225kPa 7:29
136.5 C 225kPa 7:59
136.4 C 225kPa 8:29
136.4 C 225kPa 8:29
VENTING 8:29
AIR DRYING 9:26
CYCLE COMPLETE 24:26

Digital Signature #
6B8BCE152D668437
