
Midmark M11 - v1.0.0
Total Cycles: 1929

Name: RITTER 1
Sterilizer ID: V2178198
05/11/2022 09:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.7 F 0.0
2:00 190.7 F 3.3
4:00 191.4 F 7.7
6:00 206.8 F 12.6
8:00 223.2 F 6.9
10:00 233.9 F 12.0
12:00 234.9 F 8.9
14:00 240.7 F 9.9
16:00 246.5 F 13.3
18:00 258.6 F 19.8
20:00 269.3 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.5 F 28.9
2:00 272.3 F 28.5
3:00 271.4 F 28.0
4:00 271.8 F 28.6

Min 271.2 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 209.0 F 0.0
10:00 208.3 F 0.0
15:00 209.9 F 0.0
20:00 233.0 F 0.0
25:00 239.4 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 21:08
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:26
CYCLE COMPLETE

05/11/2022 10:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____