

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001689  
12:05 24/01/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:26  
PRESSURIZING 2:32  
STERILIZING 3:42  
136.4 C 225kPa 3:42  
136.2 C 225kPa 4:12  
136.4 C 224kPa 4:42  
136.4 C 225kPa 5:12  
136.4 C 223kPa 5:42  
136.4 C 223kPa 6:12  
136.2 C 224kPa 6:42  
136.3 C 225kPa 7:12  
136.3 C 225kPa 7:12  
VENTING 7:12  
AIR DRYING 8:04  
CYCLE COMPLETE 8:20  
-----

Digital Signature #  
F2EB3B5F1C696472  
-----