
Midmark M11 - v1.0.5
Total Cycles: 2875

Name: RITTER2
Sterilizer ID: V2313228
09/13/2023 08:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 70.8 F 0.0
4:00 75.0 F 0.0
6:00 105.2 F 0.0
8:00 171.2 F 0.0
10:00 189.7 F 1.6
12:00 199.1 F 5.0
14:00 216.9 F 8.4
16:00 226.5 F 11.9
18:00 224.3 F 6.5
20:00 236.1 F 11.0
22:00 233.2 F 7.6
24:00 244.4 F 13.4
26:00 242.7 F 12.1
28:00 254.5 F 18.1
30:00 264.8 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.9
1:00 272.7 F 29.6
2:00 271.6 F 28.8
3:00 272.8 F 29.4
4:00 271.5 F 28.9

Min 271.4 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.8 F 0.0
5:00 212.9 F 0.0
10:00 186.8 F 0.0
15:00 187.3 F 0.0
20:00 197.9 F 0.0
25:00 206.4 F 0.0
30:00 209.8 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:50
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:10:02
CYCLE COMPLETE

09/13/2023 09:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____