
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003455
14:33 14/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:45
PRESSURIZING 1:51
STERILIZING 2:45
136.5 C 222kPa 2:45
136.1 C 222kPa 3:15
136.2 C 223kPa 3:45
136.2 C 222kPa 4:15
136.0 C 217kPa 4:45
136.4 C 222kPa 5:15
136.4 C 223kPa 5:45
136.4 C 222kPa 6:15
136.4 C 222kPa 6:15
VENTING 6:15
AIR DRYING 7:42
CYCLE COMPLETE 16:32

Digital Signature #
2BEC38DCC7330AE1
