

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000878  
11:25 20/01/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:22  
PRESSURIZING 3:26  
STERILIZING 4:52  
136.4 C 232kPa 4:52  
136.3 C 232kPa 5:22  
136.4 C 233kPa 5:52  
136.4 C 231kPa 6:22  
136.4 C 232kPa 6:52  
136.4 C 233kPa 7:22  
136.2 C 231kPa 7:52  
136.4 C 233kPa 8:22  
136.4 C 233kPa 8:22  
VENTING 8:22  
AIR DRYING 9:17  
CYCLE COMPLETE 16:02  
-----

Digital Signature #  
BAA9E6B07C28914D  
-----