

-----  
Midmark M11 - v1.0.5  
Total Cycles: 433

Name: RITTER2  
Sterilizer ID: V2313228  
10/08/2021 09:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.2 F 0.0  
2:00 175.5 F 3.1  
4:00 198.5 F 8.2  
6:00 218.3 F 10.4  
8:00 232.7 F 9.4  
10:00 235.5 F 9.6  
12:00 243.2 F 12.9  
14:00 242.7 F 11.9  
16:00 255.6 F 18.7  
18:00 267.0 F 25.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 272.5 F 29.3  
2:00 272.6 F 29.8  
3:00 272.0 F 28.9  
4:00 272.7 F 29.6  
  
Min 271.5 F 28.8  
Max 273.0 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.0 F 0.0  
5:00 211.0 F 0.0  
10:00 211.4 F 0.0  
15:00 200.6 F 0.0  
20:00 228.3 F 0.0  
25:00 227.1 F 0.0  
30:00 228.6 F 0.0  
Drying Complete

FILLING: 2:20  
HEATING: 19:25  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:54  
CYCLE COMPLETE

10/08/2021 10:51 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_