
Midmark M11 - v1.0.5
Total Cycles: 3081

Name: RITTER2
Sterilizer ID: V2313228
11/03/2023 10:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.6 F 0.0
2:00 171.3 F 1.6
4:00 175.7 F 5.7
6:00 201.4 F 11.1
8:00 219.1 F 5.4
10:00 231.5 F 10.9
12:00 232.9 F 8.2
14:00 240.5 F 10.7
16:00 247.9 F 14.3
18:00 261.0 F 21.6
20:00 272.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.6 F 29.1
2:00 271.5 F 28.8
3:00 272.7 F 29.1
4:00 271.5 F 28.5

Min 271.4 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.3 F 0.0
5:00 222.9 F 0.0
10:00 206.2 F 0.0
15:00 200.9 F 0.0
20:00 197.8 F 0.0
25:00 195.1 F 0.0
30:00 193.0 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 20:23
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:55
CYCLE COMPLETE

11/03/2023 11:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____