
Midmark M11 - v1.0.0
Total Cycles: 3840

Name: RITTER 1
Sterilizer ID: V2178198
09/29/2023 10:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.5 F 0.0
2:00 185.6 F 2.5
4:00 188.0 F 6.5
6:00 220.0 F 11.8
8:00 222.7 F 5.9
10:00 232.0 F 10.9
12:00 234.0 F 7.3
14:00 245.3 F 13.4
16:00 243.8 F 11.9
18:00 255.9 F 18.2
20:00 266.6 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.7 F 28.9
2:00 272.2 F 28.9
3:00 271.7 F 28.4
4:00 272.2 F 28.6

Min 270.9 F 27.9
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 213.4 F 0.0
10:00 233.1 F 0.0
15:00 228.4 F 0.0
20:00 224.3 F 0.0
25:00 221.9 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 21:30
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:00:12
CYCLE COMPLETE

09/29/2023 11:27 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____