
Midmark M11 - v1.0.5
Total Cycles: 2833

Name: RITTER2
Sterilizer ID: V2313228
09/05/2023 03:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.0 F 0.0
2:00 176.9 F 3.7
4:00 187.8 F 8.4
6:00 213.2 F 9.5
8:00 227.5 F 9.3
10:00 233.9 F 9.3
12:00 243.2 F 12.9
14:00 242.8 F 12.1
16:00 255.4 F 18.8
18:00 266.9 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.6 F 29.6
2:00 271.4 F 28.9
3:00 272.6 F 29.8
4:00 272.0 F 29.1

Min 271.3 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 214.0 F 0.0
10:00 234.4 F 0.0
15:00 234.6 F 0.0
20:00 221.9 F 0.0
25:00 219.9 F 0.0
30:00 223.5 F 0.0

Drying Complete

FILLING: 1:50
HEATING: 19:28
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:34
CYCLE COMPLETE

09/05/2023 04:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____