
Midmark M11 - v1.0.0
Total Cycles: 4055

Name: RITTER 1
Sterilizer ID: V2178198
11/23/2023 02:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 89.3 F 0.0
2:00 91.4 F 0.4
4:00 109.4 F 2.1
6:00 133.2 F 4.6
8:00 178.0 F 7.7
10:00 190.7 F 12.3
12:00 206.3 F 3.9
14:00 216.2 F 7.4
16:00 227.3 F 11.3
18:00 228.3 F 6.8
20:00 240.8 F 12.1
22:00 238.5 F 9.9
24:00 250.8 F 15.8
26:00 261.5 F 21.9
28:00 271.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.4 F 29.2
2:00 271.6 F 28.4
3:00 272.1 F 29.1
4:00 272.4 F 29.0

Min 271.2 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 211.1 F 0.0
10:00 211.4 F 0.0
15:00 197.7 F 0.0
20:00 209.4 F 0.0
25:00 226.0 F 0.0
30:00 230.6 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 28:38
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:07:16
CYCLE COMPLETE

11/23/2023 03:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____