

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 007205  
12:12 15/04/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:34  
PRESSURIZING 3:34  
STERILIZING 4:35  
136.4 C 220kPa 4:35  
136.3 C 224kPa 5:05  
136.4 C 225kPa 5:35  
136.4 C 223kPa 6:05  
136.4 C 225kPa 6:35  
136.4 C 223kPa 7:05  
136.3 C 223kPa 7:35  
136.4 C 222kPa 8:05  
136.4 C 222kPa 8:05  
VENTING 8:05  
AIR DRYING 8:58  
CYCLE COMPLETE 10:39

-----  
Digital Signature #  
DEAF35A4CF8A70DB  
-----