
Midmark M11 - v1.0.5
Total Cycles: 2731

Name: RITTER2
Sterilizer ID: V2313228
08/03/2023 03:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.5 F 0.0
2:00 175.2 F 2.4
4:00 181.9 F 6.2
6:00 202.9 F 10.8
8:00 218.6 F 5.1
10:00 229.8 F 10.3
12:00 232.9 F 8.0
14:00 243.6 F 12.4
16:00 244.4 F 12.9
18:00 256.6 F 19.6
20:00 267.7 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.6 F 29.9
2:00 271.9 F 29.1
3:00 271.9 F 29.6
4:00 272.9 F 29.8

Min 271.4 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.4 F 0.0
5:00 222.0 F 0.0
10:00 237.0 F 0.0
15:00 220.8 F 0.0
20:00 218.9 F 0.0
25:00 221.0 F 0.0
30:00 215.6 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 21:24
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:07
CYCLE COMPLETE

08/03/2023 04:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____