
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001697
8:19 08/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:50
PRESSURIZING 3:50
STERILIZING 4:53
136.0 C 220kPa 4:53
136.4 C 221kPa 5:23
136.3 C 222kPa 5:53
136.4 C 223kPa 6:23
136.4 C 222kPa 6:53
136.2 C 223kPa 7:23
135.9 C 221kPa 7:53
136.4 C 221kPa 8:23
136.4 C 221kPa 8:23
VENTING 8:23
AIR DRYING 9:18
CYCLE COMPLETE 9:29

Digital Signature #
E1332ECA77F21DB8
