

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000830  
15:27 21/12/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:58  
PRESSURIZING 3:05  
STERILIZING 4:12  
136.0 C 226kPa 4:12  
135.9 C 228kPa 4:42  
136.4 C 230kPa 5:12  
136.4 C 233kPa 5:42  
136.2 C 230kPa 6:12  
136.2 C 231kPa 6:42  
136.4 C 232kPa 7:12  
136.1 C 229kPa 7:42  
136.1 C 229kPa 7:42  
VENTING 7:42  
AIR DRYING 8:33  
CYCLE COMPLETE 16:41  
-----

Digital Signature #  
F1A24DB938D229F2  
-----