

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1101

Name: RITTER 1  
Sterilizer ID: V2178198  
08/19/2021 12:35 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.6 F 0.0  
2:00 187.2 F 4.1  
4:00 212.9 F 10.0  
6:00 222.5 F 5.5  
8:00 241.3 F 12.1  
10:00 238.6 F 10.2  
12:00 236.9 F 9.0  
14:00 252.7 F 16.5  
16:00 265.3 F 23.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.2  
1:00 271.7 F 28.1  
2:00 272.0 F 28.2  
3:00 272.3 F 28.5  
4:00 272.4 F 28.8  
  
Min 271.4 F 28.1  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.1 F 0.0  
5:00 192.4 F 0.0  
10:00 192.1 F 0.0  
15:00 191.2 F 0.0  
20:00 188.2 F 0.0  
25:00 185.4 F 0.0  
30:00 184.1 F 0.0  
Drying Complete

FILLING: 1:53  
HEATING: 17:36  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:56:38  
CYCLE COMPLETE

08/19/2021 01:50 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02