
Midmark M11 - v1.0.5
Total Cycles: 2469

Name: RITTER2
Sterilizer ID: V2313228
06/06/2023 09:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.1 F 0.0
2:00 164.0 F 2.5
4:00 167.9 F 5.8
6:00 192.0 F 10.2
8:00 213.1 F 5.5
10:00 225.4 F 9.9
12:00 230.6 F 8.4
14:00 240.7 F 12.7
16:00 240.4 F 11.3
18:00 253.4 F 17.7
20:00 264.6 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.2
1:00 271.7 F 29.7
2:00 272.6 F 29.8
3:00 271.4 F 29.1
4:00 271.8 F 29.8

Min 271.1 F 29.0
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.9 F 0.0
5:00 214.6 F 0.0
10:00 242.0 F 0.0
15:00 225.6 F 0.0
20:00 225.6 F 0.0
25:00 227.2 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 21:45
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:03
CYCLE COMPLETE

06/06/2023 10:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____