
Midmark M11 - v1.0.0
Total Cycles: 4556

Name: RITTER 1
Sterilizer ID: V2178198
05/15/2024 10:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.1 F 0.0
2:00 190.6 F 3.5
4:00 193.9 F 8.2
6:00 215.6 F 13.3
8:00 223.6 F 8.2
10:00 234.9 F 12.5
12:00 237.9 F 11.0
14:00 236.4 F 9.3
16:00 250.5 F 16.0
18:00 262.3 F 22.6
20:00 272.4 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.4 F 29.3
2:00 271.6 F 28.6
3:00 271.5 F 28.9
4:00 272.3 F 29.2

Min 271.0 F 28.3
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 219.4 F 0.0
10:00 227.0 F 0.0
15:00 230.5 F 0.0
20:00 229.4 F 0.0
25:00 227.2 F 0.0
30:00 229.7 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 20:20
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 00:59:33
CYCLE COMPLETE

05/15/2024 11:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____