
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001359
8:36 22/03/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:48
PRESSURIZING 3:50
STERILIZING 4:56
136.4 C 222kPa 4:56
136.3 C 221kPa 5:26
136.4 C 222kPa 5:56
136.4 C 222kPa 6:26
136.2 C 222kPa 6:56
136.2 C 220kPa 7:26
136.4 C 223kPa 7:56
136.3 C 224kPa 8:26
136.3 C 224kPa 8:26
VENTING 8:26
AIR DRYING 9:17
CYCLE COMPLETE 9:28

Digital Signature #
6A88484B5CAEE588
