
Midmark M11 - v1.0.0
Total Cycles: 687

Name: MIDMARK
Sterilizer ID: V2178198
02/22/2021 02:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.9 F 0.0
2:00 122.6 F 0.1
4:00 131.0 F 0.2
6:00 168.8 F 2.9
8:00 186.0 F 7.4
10:00 207.9 F 12.3
12:00 225.9 F 6.5
14:00 233.4 F 11.6
16:00 236.5 F 8.8
18:00 238.4 F 9.7
20:00 247.6 F 13.9
22:00 260.4 F 20.9
24:00 272.0 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 271.7 F 28.1
2:00 272.1 F 28.4
3:00 272.4 F 28.8
4:00 272.0 F 28.8

Min 270.9 F 28.0
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.1 F 0.0
5:00 202.8 F 0.0
10:00 214.8 F 0.0
15:00 227.1 F 0.0
20:00 233.9 F 0.0
25:00 228.5 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 24:26
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:02:22
CYCLE COMPLETE

02/22/2021 03:10 PM
ACCEPTED: X
REJECTED:
Operator: U01