
Midmark M11 - v1.0.5
Total Cycles: 2871

Name: RITTER2
Sterilizer ID: V2313228
09/12/2023 01:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.3 F 0.0
2:00 170.6 F 3.6
4:00 191.3 F 8.1
6:00 210.6 F 13.3
8:00 229.9 F 8.7
10:00 238.0 F 10.7
12:00 238.8 F 12.1
14:00 240.5 F 10.9
16:00 253.5 F 17.5
18:00 264.7 F 24.4
20:00 270.6 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.5
1:00 272.8 F 29.7
2:00 272.4 F 30.1
3:00 272.1 F 29.5
4:00 273.1 F 30.1

Min 271.5 F 29.3
Max 273.2 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.9 F 0.0
5:00 241.6 F 0.0
10:00 214.7 F 0.0
15:00 227.1 F 0.0
20:00 191.8 F 0.0
25:00 179.8 F 0.0
30:00 179.3 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 20:43
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:00:01
CYCLE COMPLETE

09/12/2023 02:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____