

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3178

Name: RITTER2  
Sterilizer ID: V2313228  
12/04/2023 09:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.4 F 0.0  
2:00 71.5 F 0.0  
4:00 79.0 F 0.2  
6:00 93.6 F 0.8  
8:00 127.0 F 3.0  
10:00 160.6 F 6.1  
12:00 177.5 F 9.5  
14:00 195.4 F 13.1  
16:00 212.9 F 5.7  
18:00 223.7 F 9.4  
20:00 231.5 F 12.9  
22:00 235.0 F 10.2  
24:00 236.1 F 9.0  
26:00 245.9 F 14.0  
28:00 257.1 F 20.0  
30:00 267.2 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.3  
1:00 272.6 F 29.9  
2:00 271.2 F 29.3  
3:00 272.3 F 30.0  
4:00 271.5 F 29.3

Min 271.2 F 29.2  
Max 272.8 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.7 F 0.0  
5:00 210.8 F 0.0  
10:00 201.1 F 0.0  
15:00 197.2 F 0.0  
20:00 191.3 F 0.0  
25:00 187.5 F 0.0  
30:00 192.1 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 31:31  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:09:47  
CYCLE COMPLETE

12/04/2023 11:08 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_