
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.2uS / 0.7ppm
CYCLE NUMBER 000606
13:42 26/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:34
STERILIZING 4:01
136.5 C 231kPa 4:01
136.4 C 233kPa 4:31
136.4 C 233kPa 5:01
136.4 C 233kPa 5:31
136.4 C 233kPa 6:01
136.2 C 231kPa 6:31
136.2 C 233kPa 7:01
136.2 C 230kPa 7:31
136.2 C 230kPa 7:31
VENTING 7:31
AIR DRYING 8:24
CYCLE COMPLETE 23:25

Digital Signature #
4006F59CE8F88687
